

Registration of 2D intraoral radiographs to 3D CBCT using modality-specific encoding

Josué Ruano-Balseca¹, Shojaat Joodi Bigdilo¹, Pieter Van Leemput², Jan De Beenhouwer¹, and Jan Sijbers¹
¹imec-Vision Lab, Department of Physics, Universiteit Antwerpen, Antwerp, Belgium, ²Envista-Medicim, Mechelen, Belgium

WHY THIS MATTERS


Preoperative Cone Beam CT (CBCT) provides the 3D anatomy for diagnosis and implant planning.

Postoperative Intraoral radiograph (IOR) provides high-resolution 2D information, revealing implant position and anatomical changes.

Implant unsafe positioning is highly prevalent (80%) [1]

OUR PROPOSAL

Generate a 2D digitally reconstructed radiograph (DRR) in the same IOR imaging pose.



THE CHALLENGE

IOR-CBCT registration is different from conventional 2D/3D registration:

- Detector is inside the oral cavity.
- Strong cross-modality appearance differences.
- Frequent metal artifacts.
- Anatomy may change between CBCT/IOR.

No paired IOR-CBCT dataset with known poses.


DATASET

839 CBCTs [2,3,4]
167,800 DRRs for training/internal evaluation (Int.).


100 CBCTs [5] → external evaluation (Ext.).

2,050 unpaired IORs
self-supervised 2D pretraining/adaptation

41 paired IOR-CBCT cases → evaluation.
These cases include metal artifacts, implants, and anatomical changes between acquisitions.



METHODOLOGY




RESULTS

Ablation study: effect of 2D/3D feature branches, ssl, and acquisition parameters (Ap.).

Comparison against baselines on a test partition from the training database (Int.) and an external dataset (Ext.).

Method	DB.	Source		Detector		Rotation
		TRE(mm)↓	SR(%)↑	TRE(mm)↓	SR(%)↑	MAE(degrees) ↓
2D	Int.	38.8	65.3	5.7	40.6	1.5
2D(ssl)	Int.	38.7	64.6	5.8	40.8	1.5
2D(ssl)+Ap.	Int.	38.6	65.7	5.6	42.3	1.5
Proposed approach	Int.	38.2	65.8	3.4	71.9	1.4
2D(ssl)+3D(ssl)+Ap.	Ext.	40.6	63.0	5.7	40.86	1.9
CMAE+MI	Int.	356	0	43.6	0.0	11.7
CMAE+NCC	Int.	355	0	38.7	0.24	11.6
ResNet50	Int.	42.8	59.6	6.7	30.9	2.4
	Ext.	47.3	53.8	9.5	13.6	3.0
Cross-Vit	Int.	47.8	53.2	8.6	17.8	3.3
	Ext.	52.0	47.1	11.5	8.2	3.9
Foundational model (ViT-Large)	Int.	43.1	60.6	6.6	31.5	2.7
	Ext.	48.0	53.0	9.5	13.0	3.2

On paired IOR-CBCT data, our method outperformed the best baseline, reducing source error from 192.3 to 119.1 mm, detector error from 34.5 to 20.3 mm, and rotation error from 8.1° to 6.1°.



TAKEAWAYS

Clinicians may benefit from this approach to visually assess, or automatically detect using AI, anatomical changes in IORs relative to stable CBCT anatomy.

REFERENCES

- B. R. Ribas et al., "Positioning errors of dental implants and their associations with adjacent structures and anatomical variations." Imaging Sci Dent, 2020.
- Altukroni, et al., "Detection of the pathological exposure of pulp using an artificial intelligence tool: a multicentric study over periapical radiographs." BMC Oral Health, 2023.
- Huang, Y., et al., "A multimodal dental dataset facilitating machine learning research and clinic services." Scientific Data, 2024.
- Li, X., "3D multimodal dental dataset based on CBCT and oral scan." Figshare, 2024.
- Rasnayaka, S., et al., "Denpar: Annotated intra-oral periapical radiographs dataset for machine learning." Scientific Data, 2025.
- Wang, C., et al., "MMDental - a multimodal dataset of tooth CBCT images with expert medical records." Scientific Data, 2025.
- Wang, Y., et al., "A multi-modal dental dataset for semi-supervised deep learning image segmentation." Scientific Data, 2025.